

Product Information

VACUETTE® Crossmatch

General Information

Application

VACUETTE® Crossmatch Tubes are used for determinations in serum / EDTA whole blood for crossmatch testing.

Materials

VACUETTE® Crossmatch Tubes are made out of PET (polyethylenetherephtalate). The caps are produced from PE (polyethylene) and are coloured by pigments purified in accordance with EN 71/3. The rubber component of the cap is composed of Brom Butyl Caoutchouc. The stabilisation ring is composed of PP (polypropylene).

Manufactured by

Greiner Bio-One GmbH, Austria

Certified according to SN EN ISO 9001:2000 and SN EN ISO 13485:2003 + AC 2007

Sterility

Sterile interior: SAL 10^{-6} (SAL = Sterility Assurance Level)

Standards: EN 552, EN 556, ANSI/AAMI/ISO 11137

Conformity

ISO 6710 "Single-use containers for venous blood specimen collection"

H1-A5 "Evacuated Tubes and Additives for Blood Specimen Collection- 5th Edition"; Approved Standard

CE mark

IVDD 98/79/EC "Directive 98/79/EC of the European Parliament and council of 27th of October 1998 on in vitro diagnostic medical devices" Classification: Other device (all devices except Annex II and self-testing devices).

Packaging

Tubes are packed in 50 pieces per rack wrapped with a polyethylene foil. Tubes are packed in 24 racks of 50 tubes. Each box is filled with 1200 tubes.

Shelf life

18 months

Product Information

VACUETTE® Crossmatch

General Information



	Labelling	Tube Label	Rack Label	Outer Carton
Manufacturers logo		X	X	X
VACUETTE® logo		X	X	X
Item number		X	X	X
Lot number		X	X	X
Expiry date		X	X	X
Fill volume		X	X	X
Fill line		X		
Tube dimension			X	
CE mark		X	X	X
Sterile R mark		X	X	X
Single use sign		X	X	X
Sterility indicator				X
Additive description		X	X	X
Cap description				X
Packaging information			X	X
Storage guidelines			X	X
Barcode label		(X)	X	X

Wear gloves during venipuncture and when handling blood collection tubes to minimise exposure hazard.